



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/755,428	01/12/2004	Roberto Pedros	2331CON2 (203-2539CON2)	9696
7590 Chief Patent Counsel U. S. Surgical, A Division of Tyco Healthcare Group, LP 150 Glover Avenue Norwalk, CT 06856			EXAMINER GETTMAN, CHRISTINA DANIELLE	
			ART UNIT 3734	PAPER NUMBER
			MAIL DATE 07/24/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/755,428

Applicant(s)

PEDROS ET AL.

Examiner

Christina D. Gettman

Art Unit

3734

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 June 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 10-19 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 10-19 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 25 August 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

Acknowledgement has been made of the Request for Continued Examination filed June 7, 2007.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) and the Intellectual Property and High Technology Technical Amendments Act of 2002 do not apply when the reference is a U.S. patent resulting directly or indirectly from an international application filed before November 29, 2000. Therefore, the prior art date of the reference is determined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

Claims 10-17 and 19 are rejected under 35 U.S.C. 102(e) as being anticipated by Ratcliff, U.S. 5,910,155, hereinafter '155.

'155 discloses in figure 32 an apparatus for substantially closing a vascular opening comprising a housing 406, at least two tissue engaging members (sides of 414), and at least two jaw members 422 or 408. It is unclear what is meant by "a shape member material" in claim 11, but the tissue engaging member or clip 414 is considered

Art Unit: 3734

a shape member material or a shape memory material because clips are typically made of nitinol, which is conductive. '155 discloses that clips 414 can be made of any appropriate biocompatible material. See col. 18, lines 35-44. The clips are a hook shaped. The apparatus also includes a deployment member 422, an elongate shaft 408 having a longitudinal slot for accommodating the tissue engaging members. The jaw members 414 includes an attachment member. The apparatus further comprises a movable actuator 422.

Claim 10 is also rejected under 35 U.S.C. 102(e) as being anticipated by Lee, U.S. 6,033,427, hereinafter Lee.

Lee discloses an apparatus for substantially closing a vascular opening comprising a housing 48, at least two tissue engaging members 94, and at least two jaw members 90. See figure 9.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 18 is rejected under 35 U.S.C. 103(a) as being unpatentable over '155 in view of Tay, U.S. 6,063,085. '155 does not expressly disclose an attachment device adapted to transmit thermal energy through application of an energy source thereto, thereby fusing the vascular tissue surrounding the vascular opening. However, Tay discloses that it is well known in the art to use thermal energy to fuse vascular tissue

Art Unit: 3734

together. See Tay, Abstract. It would have been obvious to one having ordinary skill in the art at the time of the invention to use thermal energy, as taught by Tay, in addition to clips, as taught by '155, to better seal the vascular opening.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claim 10 is rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claim 1 of U.S. Patent No. 6,248,124. Although the conflicting claims are not identical, they are not patentably distinct from each other because the '124 reference discloses all of the limitations found in claim 10 except for the addition of an electrode. Not using an electrode and merely using a housing, two tissue engaging members and two jaw members is considered obvious to one having ordinary skill in the art because a vessel opening can be closed with only those elements.

Response to Arguments

Applicant argued in the remarks filed April 9, 2007, that Ratcliff fails to teach or suggest the tissue engaging members being structured to function as required by claim 1. Applicant also argues that both Lee and Tay do not cure the deficiencies of Ratcliff.

Examiner notes that the Applicant's amendments do not require the tissue engaging members to be movable in a certain direction relative to the housing. The claims only recite that they need to *move* relative to the housing. This is taken to mean that the members need only move in some way in relation to the house. They could move up, down, outward, inward, etc. It is also noted that when reciting "being" or "adapted to," the claim language following these words is only functional, meaning a reference need only be capable of doing the same.


Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christina D. Gettman whose telephone number is 571-272-3128. The examiner can normally be reached on Monday-Friday 7:15 am to 3:45 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Hayes can be reached on 571-272-4959. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Christina Gettman
Art Unit 3734
571-272-3128


MICHAEL J. HAYES
SUPERVISORY PATENT EXAMINER